

SOLUTION CHANGE

WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 12/02/2019

OPERATOR NAME:
 KATHERINE M PADDOCK
 Permit #: 45899
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:48
INTERNAL STANDARD	VERIFIED	23:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:50
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:51
BLANK TEST	0.00	23:52

RP #013

