

SUPERVISOR TEST

 WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 03/11/2019

OPERATOR NAME:
 CORY L HURLBURT
 Permit #: 37017
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:26
INTERNAL STANDARD	VERIFIED	09:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:27
BLANK TEST	0.00	09:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:28
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:31

CA
 #257

