

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 05/20/2019

OPERATOR NAME:
ETHAN J MACURA
Permit #: 44080
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:32
INTERNAL STANDARD	VERIFIED	14:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:32
BLANK TEST	0.00	14:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:34
BLANK TEST	0.00	14:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:35
BLANK TEST	0.00	14:36

Ethán J Macura
#699

