

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 07/29/2019

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:52
INTERNAL STANDARD	VERIFIED	07:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:52
BLANK TEST	0.00	07:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:54
BLANK TEST	0.00	07:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:55
BLANK TEST	0.00	07:56

Handwritten signature and scribbles
#821

