

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/09/2019

OPERATOR NAME:
KATHERINE M PADDOCK
Permit #: 45899
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:56
INTERNAL STANDARD	VERIFIED	20:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:56
BLANK TEST	0.00	20:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:58
BLANK TEST	0.00	20:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:59
BLANK TEST	0.00	21:00

#1013

