

SUPERVISOR TEST

-----  
WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284  
-----

Date: 10/14/2019

OPERATOR NAME:  
JOSHUA T WHITNEY  
Permit #: 40382  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:55
INTERNAL STANDARD	VERIFIED	16:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:55
BLANK TEST	0.00	16:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:57
BLANK TEST	0.00	16:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:58
BLANK TEST	0.00	16:59

