

SUPERVISOR TEST

-----  
 WASHINGTON CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100284  
 -----

Date: 11/18/2019

OPERATOR NAME:  
 ROBERT J SULLIVAN  
 Permit #: 34106  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:16
INTERNAL STANDARD	VERIFIED	00:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:19
BLANK TEST	0.00	00:20

