

SUPERVISOR TEST

WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 12/16/2019

OPERATOR NAME:
 D TODD RICHE
 Permit #: 33411
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 19250

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 08:07 |
| INTERNAL STANDARD | VERIFIED | 08:08 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:08 |
| BLANK TEST | 0.00 | 08:09 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:09 |
| BLANK TEST | 0.00 | 08:10 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:10 |
| BLANK TEST | 0.00 | 08:11 |

