

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/03/2019

OPERATOR NAME:
DALE W QUESNEL
Permit #: 37510
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:01
INTERNAL STANDARD	VERIFIED	18:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:01
BLANK TEST	0.00	18:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:02
BLANK TEST	0.00	18:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:03
BLANK TEST	0.00	18:04

