

SOLUTION CHANGE

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 02/04/2019

OPERATOR NAME:  
MICHAEL O WEBER  
Permit #: 46067  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:40
INTERNAL STANDARD	VERIFIED	17:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:41
BLANK TEST	0.00	17:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:43
BLANK TEST	0.00	17:44

*Webster  
#611*

