

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 03/18/2019

OPERATOR NAME:
CORI J WINCH
Permit #: 44971
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:14
INTERNAL STANDARD	VERIFIED	22:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:16
BLANK TEST	0.00	22:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:17
BLANK TEST	0.00	22:18

C. Winch
697

