

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 09/30/2019

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:42
INTERNAL STANDARD	VERIFIED	08:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:42
BLANK TEST	0.00	08:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:44
BLANK TEST	0.00	08:45

