

SOLUTION CHANGE

-----  
 HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408  
 -----

Date: 08/02/2019

OPERATOR NAME:  
 RICHARD J ARNOLD  
 Permit #: 33860  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 190130

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:14
INTERNAL STANDARD	VERIFIED	17:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	17:15
BLANK TEST	0.00	17:16
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.10	17:16
BLANK TEST	0.00	17:17
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	17:17
BLANK TEST	0.00	17:18

