

SUPERVISOR TEST

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

ate: 09/18/2019

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:55
INTERNAL STANDARD	VERIFIED	18:55
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.10	18:56
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:57
BLANK TEST	0.00	18:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:58
BLANK TEST	0.00	18:59

